

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Dep. Johnson AGENCY LPCSO
DATE 1-14-16 TIME 313
INTOXILYZER SERIAL 80-000917 LOCATION QUARTZSITE

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0.096 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____
DPS Form Exh G-2 (Rev 05-07)

INTOXILYZER 8000
Location: QUARTZSITE PD
Serial Number: 80400917
Core Version: 810548
01/14/2016

03:13:20

Standard Lot#: 0704100A1
Last Changed By: R. ROSADO #557

QAS: M. JOHNSON #531
LA PAZ CSO

Operator: M. JOHNSON #531
LA PAZ CSO

Test	5/210L	Time
Air Blank	0.000	03:14:46
0.100 Cal Check	0.096	03:15:06
Air Blank	0.000	03:15:36

Preventative Maintenance
Performed -- Timer Reset