

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Mike Johnson AGENCY LPCSO  
DATE 6-12-86 TIME \_\_\_\_\_  
INTOXILYZER SERIAL # 80-000917 LOCATION QUARTZSITE

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.091 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE \_\_\_\_\_

INTOXILYZER 8000

Location: QUARTZSITE PD

Serial Number: 80-000917

Core Version: 810948

06/12/2016

15:43:09

Standard Lot#: 060610DA1

Last Changed By: J. BERRIE #608

QAS: M. JOHNSON #531

LA PAZ CSO

Operator: M. JOHNSON #531

LA PAZ CSO

Test	5/210L	Time
------	--------	------

Air Blank	0.000	15:43:50
-----------	-------	----------

0.100 Cal Check	0.099	15:44:11
-----------------	-------	----------

Air Blank	0.000	15:44:41
-----------	-------	----------

Preventative Maintenance  
Performed -- Timer Reset