

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Mike Johnson AGENCY CPCSO
DATE 6-12-16 TIME _____
INTOXILYZER SERIAL # 80-001031 LOCATION SALOME

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

INTOXILYZER 8000

Location: SALOME

Serial Number: 80-001031

Core Version: 8105.48

06/12/2016

16:27:20

Standard Lot#: 08915100A2

Last Changed By: J. GOMEZ #220

QAS: M. JOHNSON #531

LA PAZ CSO

Operator: M. JOHNSON #531

LA PAZ CSO

Test	g/210L	Time
Air Blank	0.000	16:28:08
0.100 Cal Check	0.099	16:28:29
Air Blank	0.000	16:28:59

Preventative Maintenance
Performed -- Timer Reset