

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Dep Johnson AGENCY LPCSO
DATE 12-17-15 TIME 1153
INTOXILYZER SERIAL 80-001031 LOCATION SALOME

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- ~~2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"~~
~~3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.~~
~~4. Air blank completed.~~
~~5. Calibration check completed. Test results 0.100 AC.~~
~~6. Air blank completed.~~
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

DPS Form Exh G-2 (Rev 05-01)

117
2/28

ENCLOSURE 0000

Location: SALOME

Serial Number: 80-001031

Core Version: 8105.48

12/17/2015

11:53:54

Standard Lot#: 08915100A2

Last Changed By: J. GOMEZ #220

OAS: M. JOHNSON #531

LA PAZ CSO

Operator: M. JOHNSON #531

LA PAZ CSO

Test	g/210L	Time
Air Blank	0.000	11:54:23
0.100 Cal Check	0.100	11:54:44
Air Blank	0.000	11:55:14

Preventative Maintenance
Performed -- Timer Reset