

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST E. Schneider AGENCY LP CSO  
DATE 4/25/15 TIME 1746  
INTOXILYZER SERIAL # 80-001196 LOCATION Salome Substation

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 096 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

INTOXILYZER 8000

Location: SALOME SUBSTATION

Serial Number: 80-001196

Core Version: 8105.48

04/25/2015

14:45:39

-----  
Standard Lot#: 08513100A1

Last Changed By: J. GOMEZ #220  
-----

QAS: E. SCHNEIDER #431

LA PAZ CSO  
-----

Operator: E. SCHNEIDER #431

LA PAZ CSO  
-----

Test	g/210L	Time
Air Blank	0.000	14:46:11
0.100 Cal Check	0.096	14:46:32
Air Blank	0.000	14:47:01

Preventative Maintenance  
Performed -- Timer Reset