

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST E. Schneider AGENCY LPCSO
DATE 9/17/15 TIME 0031
INTOXILYZER SERIAL # 80-001196 LOCATION Ehrenberg Substation

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 096 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE EJ

INTOXILYZER 8000
Location: EHRENBURG SUBSTATION
Serial Number: 80-001196
Core Version: 8105.48
09/15/2015 00:31:00

Standard Lot#: 08915100A2
Last Changed By: E. SCHNEIDER #431

QAS: E. SCHNEIDER #431
LA PAZ CSO

Operator: E. SCHNEIDER #431
LA PAZ CSO

Test	g/210L	Time
Air Blank	0.000	00:31:32
0.100 Cal Check	0.096	00:31:52
Air Blank	0.000	00:32:22

Preventative Maintenance
Performed -- Timer Reset