

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Mike Johnson AGENCY LPCSO
DATE 6-12-10 TIME 1421
INTOXILYZER SERIAL # 80-001197 LOCATION JAIL

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 101 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

INTOXILYZER 8000

Location: BACK UP/LPCSO ADF

Serial Number: 80-001197

Core Version: 8105.48

06/12/2016

14:21:12

Standard Lot#: 14014100A1

Last Changed By: R. ROSADO #557

QAS: M. JOHNSON #531

LA PAX CSO

Operator: M. JOHNSON #531

LA PAX CSO

Test	g/210L	Time
Air Blank	0.000	14:21:54
0.100 Cal Check	0.101	14:22:15
Air Blank	0.000	14:22:44

Preventative Maintenance
Performed -- Timer Reset