

Department of Public Safety - Alcohol Testing

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. ROSAIDO # 557 AGENCY LA PAZ COUNTY SO
DATE 12-23-14 21:20 TIME
INTOXILYZER SERIAL # 80-601197 (APA250) LOCATION LPCSO COUNTY PARK

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.101 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE R. ROSAIDO # 557

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000

Location: BACK UP/COUNTY PARK

Serial Number: 80-001197

Core Version: 8105.48

12/23/2014

21:22:03

Standard Lot#: 14014100A1

Last Changed By: R. ROSADO #557

QAS: R. ROSADO #557

LA PAZ CSO

Operator: R. ROSADO #557

LA PAZ CSO

Test	g/210L	Time
Air Blank	0.000	21:22:41
0.100 Cal Check	0.101	21:23:02
Air Blank	0.000	21:23:31

Preventative Maintenance
Performed -- Timer Reset

Department of Public Safety - Alcohol Testing

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. Rosado #557

AGENCY LA PAZ COUNTY SO

DATE 12-23-14


21:25

TIME

INTOXILYZER SERIAL # 80-001197 LAPAZSO

LOCATION LPCS0 COUNTY PARK

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 101 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  #557

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000

Location: BACK UP/COUNTY PARK

Serial Number: 80-001197

Core Version: 8105.48

12/23/2014

21:25:51

Standard Lot#: 14014100A1

Last Changed By: R. ROSADO #557

QAS: R. ROSADO #557

LA PAZ CSO

Operator: R. ROSADO #557

LA PAZ CSO

Test	g/210L	Time
Air Blank	0.000	21:26:20
0.100 Cal Check	0.101	21:26:41
Air Blank	0.000	21:27:10

Preventative Maintenance
Performed -- Timer Reset

Department of Public Safety - Alcohol Testing

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST R. Rosado #557 AGENCY LA PAZ COUNTY SO

DATE 12-23-14 TIME 21:27
INTOXILYZER SERIAL # 80-001197 LOCATION LPCSO COUNTY PARK

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS _____

SIGNATURE R. Rosado #557

DPS Form Exh G-4 (Rev 05-01)

Historical Note

New Exhibit G-4 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000

Location: BACK UP/COUNTY PARK

Serial Number: 80-001197

Core Version: 8105.48

12/23/2014

21:27:39

Standard Lot#: 14014100A1

Last Changed By: R. ROSADO

#557

QAS: R. ROSADO

#557

LA PAZ CSO

Operator: R. ROSADO

#557

LA PAZ CSO

Subject: SQAP, TEST,

DOB: 01/01/1990

Sex: M

Weight: 900

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:28:23
Diagnostic Test	Pass	21:28:58
Air Blank	0.000	21:29:26
0.100 Cal Check	0.101	21:29:47
Air Blank	0.000	21:30:16
Subject Test	0.000	21:30:47
Air Blank	ABT*	21:30:53
Air Blank	IPS**	21:31:01

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: BACK UP/COUNTY PARK

Serial Number: 80-001197

Core Version: 8105.48

12/23/2014

21:31:16

Standard Lot#: 14014100A1

Last Changed By: R. ROSADO

#557

QAS: R. ROSADO

#557

LA PAZ CSO

Operator: R. ROSADO

#557

LA PAZ CSO

Subject: SQAP, TEST,

DOB: 01/01/1990

Sex: M

Weight: 900

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:32:03
Diagnostic Test	Pass	21:32:38
Air Blank	0.000	21:33:05
0.100 Cal Check	0.101	21:33:26
Air Blank	0.000	21:33:56
Subject Test	0.216*	21:37:03
Air Blank	0.000	21:37:36
Air Blank	0.000	21:38:03
Subject Test	XXX**	21:38:36
Air Blank	RFI***	21:38:50

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***RFI Detect

Not a Successfully Completed Test Sequence

Department of Public Safety - Alcohol Testing

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY


STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R Rosendo #557 AGENCY LA PAZ COUNTY SO

DATE 12-23-14 TIME 21:39
INTOXILYZER SERIAL # 80-001107 LAPAZSO LOCATION LPCSO COUNTY PARK

- (~~t~~) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C
- (~~t~~) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (~~t~~) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (~~P~~) 4. Air blank completed.
- (~~L~~) 5. Calibration check completed. Test results 0. 098 AC.
- (~~t~~) 6. Air blank completed.
- (~~t~~) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000

Location: BACK UP/COUNTY PARK

Serial Number: 80-001197

Core Version: 8105.48

12/23/2014

21:39:06

Standard Lot#: 14014100A1

Last Changed By: R. ROSADO #557

QAS: R. ROSADO #557

LA PAZ CSO

Operator: R. ROSADO #557

LA PAZ CSO

Test	g/210L	Time
Air Blank	0.000	21:39:35
0.100 Cal Check	0.098	21:39:56
Air Blank	0.000	21:40:26

Preventative Maintenance
Performed -- Timer Reset