

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST DEPUTY JOHN M. GOMEZ #220

AGENCY LA PAZ COUNTY SHERIFF'S OFFICE

DATE 02 / 09 / 2013

TIME 20:41 Hrs

INTOXILYZER SERIAL # 80 - 001198

LOCATION LPCSO EHRENBERG STATION

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34 C 0.2 C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0.097 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

INTOXILYZER 8000
Location: EHRENBURG STATION
Serial Number: 80-001198
Core Version: 8105.48
02/09/2013

20:41:39

Standard Lot#: 15712100A1
Last Changed By: R. EPPS #90

QAS: J. GOMEZ #220
LA PAZ CSO

Operator: J. GOMEZ #220
LA PAZ CSO

Test	g/210L	Time
Air Blank	0.000	20:42:56
0.100 Cal Check	0.099	20:43:17
Air Blank	0.000	20:43:46

Preventative Maintenance
Performed -- Timer Reset

