

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Dep Johnson #531 AGENCY LPCSO
DATE 3-13-16 TIME 1036
INTOXILYZER SERIAL # 80-003861 LOCATION BOATHOUSE

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 100 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE _____

INTOXILYZER 8000
Location: LPCSO BOATHOUSE
Serial Number: 80-003861
Core Version: 8105.48
03/13/2016

10:37:45

Standard Lot#: 14014100A1
Last Changed By: E. SCHNEIDER #431

QAS: M. JOHNSON #531
LA PAZ CSO

Operator: M. JOHNSON #531
LA PAZ CSO

Test	g/210L	Time
Air Blank	0.000	10:38:18
0.100 Cal Check	0.100	10:38:39
Air Blank	0.000	10:39:09

Preventative Maintenance
Performed -- Timer Reset