

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST E. Schneider AGENCY LPCSO
DATE 8/23/15 TIME 0802
INTOXILYZER SERIAL # 80-003861 LOCATION Boat house

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 099 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE E. Schneider

INTOXILYZER 8000

Location: LPCSO BOATHOUSE

Serial Number: 80-003861

Core Version: 8105.48

08/23/2015

08:02:55

Standard Lot#: 14014100A1

Last Changed By: E. SCHNEIDER #431

QAS: E. SCHNEIDER #431

LA PAZ CSO

Operator: E. SCHNEIDER #431

LA PAZ CSO

Test	g/210L	Time
Air Blank	0.000	08:03:26
0.100 Cal Check	0.099	08:03:47
Air Blank	0.000	08:04:17

Preventative Maintenance
Performed -- Timer Reset