

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. ROSADO #557 AGENCY LA PAZ COUNTY SO

DATE 12-23-14 TIME 20:39  
INTOXILYZER SERIAL # 20-003861 LAPAZSO LOCATION LPCSO BOMBING SAFETY

- (  ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C
- (  ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (  ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (  ) 4. Air blank completed.
- (  ) 5. Calibration check completed. Test results 0. 0.98 AC.
- (  ) 6. Air blank completed.
- (  ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE 

DPS Form Exh G-2 (Rev 05-01)

**Historical Note**

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000

Location: LPCSO BOATHOUSE

Serial Number: 80-003861

Core Version: 8105.48

12/23/2014

20:39:01

-----  
Standard Lot#: 08513100A1

Last Changed By: J. GOMEZ #220  
-----

QAS: R. ROSADO #557

LA PAZ CSO  
-----

Operator: R. ROSADO #557

LA PAZ CSO  
-----

Test	g/210L	Time
Air Blank	0.000	20:39:51
0.100 Cal Check	0.098	20:40:12
Air Blank	0.000	20:40:41

Preventative Maintenance  
Performed -- Timer Reset

Department of Public Safety - Alcohol Testing

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. ROSADO #557 AGENCY LA PAZ COUNTY SO

DATE 12-23-11 20:40 TIME  
INTOXILYZER SERIAL # 80-003861 CAPAZSO LOCATION LPCSO BOATING SAFETY

- (  ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C
- (  ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (  ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (  ) 4. Air blank completed.
- (  ) 5. Calibration check completed. Test results 0. 098 AC.
- (  ) 6. Air blank completed.
- (  ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  #557

DPS Form Exh G-2 (Rev 05-01)

Historical Note  
New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000

Location: LPCSO BOATHOUSE

Serial Number: 80-003861

Core Version: 8105.48

12/23/2014

20:42:39

-----  
Standard Lot#: 08513100A1

Last Changed By: J. GOMEZ #220  
-----

QAS: R. ROSADO #557

LA PAZ CSO  
-----

Operator: R. ROSADO #557

LA PAZ CSO  
-----

Test	g/210L	Time
Air Blank	0.000	20:43:09
0.100 Cal Check	0.098	20:43:29
Air Blank	0.000	20:43:59

Preventative Maintenance  
Performed -- Timer Reset

Department of Public Safety - Alcohol Testing

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST R. ROSADO #557 AGENCY LA PAZ COUNTY SO

DATE 12-23-14 TIME 20:44  
INTOXILYZER SERIAL # 80-003361 LAPAZSO LOCATION LPCSO BONTING SAFETY

( 6 ) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ( 6 ) 1. Clock time check.
- ( 8 ) 2. Date check.

OPERATIONAL TESTS

- ( 6 ) 1. Alcohol-free subject test result 0. 000 AC.
- ( 8 ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- ( 6 ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- ( < ) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES X NO     

COMMENTS \_\_\_\_\_

AMBIENT FAIL ON CAL CHECK #1

SIGNATURE 

DPS Form Exh G-4 (Rev 05-01)

Historical Note

New Exhibit G-4 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).

INTOXILYZER 8000

Location: LPCSO BOATHOUSE

Serial Number: 80-003861

Core Version: 8105.48

12/23/2014

20:44:08

Standard Lot#: 08513100A1

Last Changed By: J. GOMEZ #220

QAS: R. ROSADO #557

LA PAZ CSO

Operator: R. ROSADO #557

LA PAZ CSO

Subject: SQAP, TEST,

DOB: 01/01/1990

Sex: M Weight: 900

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	20:44:56
Diagnostic Test	Pass	20:45:31
Air Blank	0.000	20:45:58
0.100 Cal Check	0.099	20:46:19
Air Blank	0.000	20:46:49
Subject Test	0.000	20:47:13
Air Blank	ABT*	20:47:21
Air Blank	IPS**	20:47:29

\*Sequence Aborted

\*\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: LPCSO BOATHOUSE  
Serial Number: 80-003861  
Core Version: 8105.48  
12/23/2014

20:47:37

---

Standard Lot#: 08513100A1  
Last Changed By: J. GOMEZ #220

---

QAS: R. ROSADO #557  
LA PAZ CSO

---

Operator: R. ROSADO #557  
LA PAZ CSO

---

Subject: SQAP,TEST,  
DOB: 01/01/1990  
Sex: M Weight: 900

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	20:48:21
Diagnostic Test	Pass	20:48:56
Air Blank	0.000	20:49:23
0.100 Cal Check	0.099	20:49:44
Air Blank	0.000	20:50:14
Subject Test	0.158*	20:53:21
Air Blank	0.000	20:53:55
Air Blank	0.000	20:54:22
Subject Test	RGE**	20:55:02
Air Blank	RFI***	20:55:12

\*Deficient Sample  
\*\*Range Exceeded  
\*\*\*RFI Detect

Not a Successfully Completed Test Sequence

Department of Public Safety - Alcohol Testing

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST R. ROSADO #557

AGENCY LA PAZ COUNTY SO

DATE 12-23-14 20:55 TIME  
INTOXILYZER SERIAL # 80-003861 LAPAZSO LOCATION LPCSO BOATING SAFETY

- (  ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34° C ± 0.2° C
- (  ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (  ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (  ) 4. Air blank completed.
- (  ) 5. Calibration check completed. Test results 0. 097 AC.
- (  ) 6. Air blank completed.
- (  ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE R. Rosado #557

DPS Form Exh G-2 (Rev 05-01)

Historical Note

New Exhibit G-2 made by final rulemaking at 12 A.A.R. 1916, effective 9:00 a.m., May 18, 2006 (Supp. 06-2).



INTOXILYZER 8000

Location: LPCSO BOATHOUSE

Serial Number: 80-003861

Core Version: 8105.48

12/23/2014

20:55:35

-----  
Standard Lot#: 08513100A1

Last Changed By: J. GOMEZ #220  
-----

QAS: R. ROSADO #557

LA PAZ CSO  
-----

Operator: R. ROSADO #557

LA PAZ CSO  
-----

Test	g/210L	Time
Air Blank	AMB*	20:56:05
Air Blank	0.000	20:56:33

\*Ambient Fail

INTOXILYZER 8000

Location: LPCSO BOATHOUSE

Serial Number: 80-003861

Core Version: 8105.48

12/23/2014

20:56:36

-----  
Standard Lot#: 08513100A1

Last Changed By: J. GOMEZ #220  
-----

QAS: R. ROSADO #557

LA PAZ CSO  
-----

Operator: R. ROSADO #557

LA PAZ CSO  
-----

Test	g/210L	Time
Air Blank	0.000	20:57:06
0.100 Cal Check	0.097	20:57:27
Air Blank	0.000	20:57:57

Preventative Maintenance  
Performed -- Timer Reset