

EXHIBIT G-2

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R-13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST DEPUTY JOHN M. GOMEZ #220

AGENCY LA PAZ COUNTY SHERIFF'S OFFICE

DATE 02 / 02 / 2013

TIME 15:51 Hrs

INTOXILYZER SERIAL # 80 - 000934

LOCATION LPCSO JAIL

(✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.

OR

Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34 C 0.2 C

(✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

(✓) 4. Air blank completed.

(✓) 5. Calibration check completed. Test results 0. 099 AC.

(✓) 6. Air blank completed.

(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE \_\_\_\_\_

INTOXILYZER 8000  
Location: LPCSO JAIL  
Serial Number: 80-000934  
Core Version: 8105.48  
02/02/2013

15:51:04

-----  
Standard Lot#: 15712100A1  
Last Changed By: R. EPPS #90  
-----

QAS: J. GOMEZ #220  
LA PAZ CSO  
-----

Operator: J. GOMEZ #220  
LA PAZ CSO  
-----

Test	g/210L	Time
Air Blank	0.000	15:52:30
0.100 Cal Check	0.097	15:52:51
Air Blank	0.000	15:53:21

Preventative Maintenance  
Performed -- Timer Reset



*[Handwritten Signature]*  
#220