

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Dep Johnson AGENCY LPCSO
DATE 12-6-14 TIME 1844
INTOXILYZER SERIAL # 80-000934 LOCATION LPCSO JAIL

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach this record to the completed checklist.

SIGNATURE [Signature]

DPS Form Exh G-2 (Rev 05-01)

MUST WAKE UP INSTRUMENT
CANT Perform in Standby Mode

INTOXILYZER 8000
Location: LPCSO JAIL
Serial Number: 80-000934
Core Version: 8105.48
12/06/2014

18:44:44

Standard Lot#: 08513100A1
Last Changed By: M. JOHNSON #531

QAS: M. JOHNSON #531
LA PAZ CSO

Operator: M. JOHNSON #531
LA PAZ CSO

Test	g/210L	Time
Air Blank	0.000	18:45:19
0.100 Cal Check	0.099	18:45:40
Air Blank	0.000	18:46:10

Preventative Maintenance
Performed -- Timer Reset